

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 925

Volumetric pipette, 250 ml, AS, conf. cert.

Batch: 06-24


Serial:

Measuring Nr. 66471

Batch size:

14

Volume:	250,000 ml
Tolerance:	$\pm 0,0800$ ml
Average value:	249,9994 ml
Deviation:	0,0186 ml
Precision (CV):	0,0074 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:

4

Date:

26.03.2024

Balance:

Sartorius - LA
1200 S

DE-M