

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 555 005

**Volumetric pipettes, 0,5 ml, 2 marks, AS, conf. cert.**


**Batch:** 11-18

**Serial:**

**Measuring Nr.** 19669

**Batch size:** 1185

<b>Volume:</b>	<b>0,5000 ml</b>
<b>Tolerance:</b>	$\pm 0,0050$ ml
<b>Average value:</b>	0,5001 ml
<b>Deviation:</b>	0,0010 ml
<b>Precision (CV):</b>	0,1942 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	44
<b>Date:</b>	10.04.2018
<b>Balance:</b>	Sartorius - LE244S-OCE

DE-M