

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 555 020

**Volumetric pipettes, 2 ml, with 2 marks, AS, conf. cert.**

**Batch:** 04-18


**Serial:**

**Measuring Nr.** 17475

**Batch size:**

1

<b>Volume:</b>	2,0000 ml
<b>Tolerance:</b>	$\pm 0,0100$ ml
<b>Measured value:</b>	1,9997 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

9

**Date:**

24.01.2018

**Balance:**

Sartorius -  
LE244S-OCE

DE-M