

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 555 020

**Volumetric pipettes, 2 ml, 2 marks, AS, conf. cert.**


**Batch:** 27-22

**Serial:**

**Measuring Nr.** 58436

**Batch size:** 500

|                        |                 |
|------------------------|-----------------|
| <b>Volume:</b>         | 2,0000 ml       |
| <b>Tolerance:</b>      | $\pm 0,0100$ ml |
| <b>Average value:</b>  | 1,9991 ml       |
| <b>Deviation:</b>      | 0,0022 ml       |
| <b>Precision (CV):</b> | 0,1122 %        |



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

|                  |                           |
|------------------|---------------------------|
| <b>Operator:</b> | 7                         |
| <b>Date:</b>     | 26.08.2022                |
| <b>Balance:</b>  | Sartorius -<br>LE244S-OCE |

DE-M