

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 555 300

Volumetric pipettes, 30 ml, 2 marks, AS, conf.cert.


Batch: 44-19

Serial:

Measuring Nr. 34589

Batch size: 150

Volume:	30,000 ml
Tolerance:	$\pm 0,0300$ ml
Average value:	30,0053 ml
Deviation:	0,0076 ml
Precision (CV):	0,0253 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	19.11.2019
Balance:	Sartorius - MSE224S-0CE -DU

DE-M