

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 555 500

**Volumetric pipettes, 50 ml, with 2 marks, AS, conf. cert.**

**Batch:** 19-22


**Serial:**

**Measuring Nr.** 56512

**Batch size:**

277

<b>Volume:</b>	50,0000 ml
<b>Tolerance:</b>	$\pm 0,0500$ ml
<b>Average value:</b>	50,0003 ml
<b>Deviation:</b>	0,0076 ml
<b>Precision (CV):</b>	0,0153 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

7

**Date:**

08.06.2022

**Balance:**

Sartorius -  
LE244S-OCE

DE-M