

# Certificate of performance



**Article:** 3 578 710 BL

**Grad. pipettes, 10:0,1 ml, blue grad., AS, conf. cert.**




**Batch:** 34-17

**Serial:**

**Measuring Nr.** 14058

**Batch size:** 60

Volume:	10,000 ml	6,000 ml	2,000 ml
<b>Tolerance:</b>	± 0,0500 ml	± 0,0500 ml	± 0,0500 ml
<b>Average value:</b>	9,9972 ml	5,9999 ml	2,0033 ml
<b>Deviation:</b>	0,0094 ml	0,0096 ml	0,0132 ml
<b>Precision (CV):</b>	0,0939 %	0,1593 %	0,6592 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 835

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	9
<b>Date:</b>	12.09.2017
<b>Balance:</b>	Sartorius - LE244S-OCE

