

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 124 256

Pellet-burettes, 25 ml, amber glass, AS, conf. cert.




Batch: 44-21

Serial:

Measuring Nr.: 65422

Batch size: 6

Volume:	25,000 ml	15,000 ml	5,000 ml
Tolerance:	± 0,0300 ml	± 0,0300 ml	± 0,0300 ml
Average value:	25,0123 ml	15,0144 ml	5,0174 ml
Deviation:	0,0070 ml	0,0042 ml	0,0051 ml
Precision (CV):	0,0281 %	0,0281 %	0,1020 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 385

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	05.12.2023
Balance:	Sartorius - LE244S-OCE

