

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 005 USP

**Volumetric pipettes, 0,5 ml, USP, AS, conf. cert.**

**Batch:** 38-24


**Serial:**

**Measuring Nr.** 69868

**Batch size:**

147

<b>Volume:</b>	0,5000 ml
<b>Tolerance:</b>	$\pm 0,0060$ ml
<b>Average value:</b>	0,5002 ml
<b>Deviation:</b>	0,0007 ml
<b>Precision (CV):</b>	0,1478 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	4
<b>Date:</b>	29.10.2024
<b>Balance:</b>	Sartorius - MSE224S-0CE -DU

DE-M